
Midmark M11 - v1.0.0
Total Cycles: 2451

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 195.6 F 3.9
4:00 213.6 F 9.3
6:00 222.2 F 6.2
8:00 233.2 F 10.0
10:00 237.5 F 8.4
12:00 243.8 F 13.2
14:00 242.8 F 12.0
16:00 256.0 F 18.8
18:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.5 F 29.2
2:00 271.6 F 28.6
3:00 272.7 F 29.3
4:00 272.0 F 28.7

Min 271.6 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 230.1 F 0.0
10:00 232.7 F 0.0
15:00 229.4 F 0.0
20:00 224.6 F 0.0
25:00 223.7 F 0.0
Dry Cycle Aborted

FILLING: 2:51
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 28:20
TOTAL CYCLE: 00:57:49

10/18/2022 02:47 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____