

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3365

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2024 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.6 F 0.0  
2:00 155.4 F 2.2  
4:00 171.9 F 5.9  
6:00 195.6 F 10.3  
8:00 213.8 F 5.0  
10:00 223.4 F 9.2  
12:00 234.5 F 10.2  
14:00 242.4 F 12.3  
16:00 242.0 F 11.4  
18:00 255.0 F 18.2  
20:00 266.5 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.6 F 29.7  
2:00 272.4 F 29.7  
3:00 271.7 F 29.4  
4:00 271.9 F 29.1

Min 270.9 F 28.5  
Max 273.1 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 198.3 F 0.0  
10:00 199.6 F 0.0  
15:00 198.9 F 0.0  
20:00 196.0 F 0.0  
25:00 193.6 F 0.0  
30:00 193.1 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:60  
CYCLE COMPLETE

02/07/2024 03:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_