

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2795

Name: RITTER2  
Sterilizer ID: V2313228  
08/25/2023 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.6 F 0.0  
2:00 76.5 F 0.0  
4:00 81.5 F 0.0  
6:00 96.0 F 0.0  
8:00 153.9 F 0.0  
10:00 192.1 F 0.2  
12:00 202.4 F 3.1  
14:00 216.6 F 6.0  
16:00 224.3 F 9.1  
18:00 233.6 F 12.8  
20:00 231.4 F 8.1  
22:00 242.3 F 12.7  
24:00 239.4 F 10.6  
26:00 236.3 F 8.9  
28:00 249.9 F 15.4  
30:00 260.9 F 21.7  
32:00 271.0 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.1  
2:00 272.0 F 29.2  
3:00 272.1 F 28.8  
4:00 272.3 F 29.5  
  
Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.6 F 0.0  
5:00 232.4 F 0.0  
10:00 229.2 F 0.0  
15:00 226.7 F 0.0  
20:00 222.3 F 0.0  
25:00 214.0 F 0.0  
30:00 212.7 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:42  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:10:47  
CYCLE COMPLETE

08/25/2023 09:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_