
Midmark M11 - v1.0.5
Total Cycles: 2020

Name: RITTER2
Sterilizer ID: V2313228
02/07/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.7 F 0.0
2:00 159.8 F 0.1
4:00 181.3 F 2.3
6:00 203.1 F 6.6
8:00 220.0 F 11.7
10:00 227.1 F 7.3
12:00 240.5 F 13.0
14:00 240.7 F 11.6
16:00 240.6 F 11.0
18:00 254.6 F 18.2
20:00 266.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.3 F 29.3
2:00 271.9 F 29.4
3:00 272.3 F 29.3
4:00 271.7 F 29.3

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 222.6 F 0.0
10:00 245.7 F 0.0
15:00 223.1 F 0.0
20:00 221.5 F 0.0
25:00 225.0 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

02/07/2023 03:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____