

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001217  
16:11 17/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:50  
STERILIZING 4:00  
136.3 C 222kPa 4:00  
136.3 C 224kPa 4:30  
136.0 C 222kPa 5:00  
136.3 C 225kPa 5:30  
136.3 C 224kPa 6:00  
136.4 C 225kPa 6:30  
136.4 C 225kPa 7:00  
136.2 C 224kPa 7:30  
136.2 C 224kPa 7:30  
VENTING 7:30  
AIR DRYING 8:22  
CYCLE COMPLETE 10:42  
-----

Digital Signature #  
AA89745F63695E0D  
-----