
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002422
12:11 14/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:47
STERILIZING 3:59
136.4 C 224kPa 3:59
136.2 C 223kPa 4:29
136.2 C 225kPa 4:59
136.3 C 225kPa 5:29
136.4 C 225kPa 5:59
136.4 C 225kPa 6:29
136.4 C 225kPa 6:59
136.4 C 225kPa 7:29
136.4 C 225kPa 7:29
VENTING 7:29
AIR DRYING 8:27
CYCLE COMPLETE 8:52

Digital Signature #
6F54AF4222E2E9F7
