

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2971

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2023 08:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.3 F 0.0  
2:00 77.2 F 0.0  
4:00 90.4 F 0.0  
6:00 111.8 F 0.0  
8:00 167.1 F 0.0  
10:00 187.0 F 1.7  
12:00 195.5 F 5.0  
14:00 210.0 F 8.4  
16:00 217.6 F 11.8  
18:00 225.4 F 5.8  
20:00 234.6 F 10.2  
22:00 236.5 F 9.0  
24:00 243.5 F 12.3  
26:00 240.8 F 10.4  
28:00 253.1 F 16.6  
30:00 264.1 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 271.6 F 28.0  
2:00 272.1 F 28.3  
3:00 272.4 F 28.6  
4:00 271.9 F 28.6  
  
Min 271.3 F 27.8  
Max 272.4 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.2 F 0.0  
5:00 217.7 F 0.0  
10:00 220.7 F 0.0  
15:00 226.3 F 0.0  
20:00 225.1 F 0.0  
25:00 223.1 F 0.0  
30:00 223.8 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:58  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:09:60  
CYCLE COMPLETE

03/06/2023 09:27 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_