

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1060

Name: RITTER 1  
Sterilizer ID: V2178198  
07/27/2021 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.0 F 0.0  
2:00 113.6 F 0.1  
4:00 127.9 F 1.2  
6:00 161.6 F 3.8  
8:00 187.7 F 6.9  
10:00 207.0 F 11.4  
12:00 219.0 F 4.9  
14:00 228.4 F 9.6  
16:00 232.4 F 9.8  
18:00 240.8 F 11.3  
20:00 237.6 F 9.2  
22:00 250.9 F 15.6  
24:00 262.1 F 22.0  
26:00 272.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 272.0 F 28.7  
2:00 272.5 F 28.6  
3:00 271.1 F 28.1  
4:00 272.6 F 29.0

Min 271.1 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 239.6 F 0.0  
10:00 217.8 F 0.0  
Dry Cycle Aborted

FILLING: 0:58  
HEATING: 26:21  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 12:34  
TOTAL CYCLE: 00:47:13

07/27/2021 01:06 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01