

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4968

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2024 03:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 169.5 F 1.5  
4:00 202.2 F 5.9  
6:00 220.7 F 11.1  
8:00 224.3 F 5.1  
10:00 236.6 F 10.2  
12:00 234.7 F 8.0  
14:00 245.7 F 13.0  
16:00 243.3 F 11.7  
18:00 256.2 F 18.3  
20:00 267.4 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.7 F 28.8  
2:00 272.1 F 28.1  
3:00 271.5 F 28.3  
4:00 272.7 F 28.8  
  
Min 271.2 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 224.8 F 0.0  
10:00 220.5 F 0.0  
15:00 226.3 F 0.0  
20:00 185.7 F 0.0  
25:00 185.1 F 0.0  
30:00 186.3 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:00:16  
CYCLE COMPLETE

10/04/2024 04:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_