
Midmark M11 - v1.0.5
Total Cycles: 2717

Name: RITTER2
Sterilizer ID: V2313228
08/01/2023 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.0 F 0.0
2:00 197.5 F 1.5
4:00 214.3 F 7.4
6:00 225.8 F 13.2
8:00 233.9 F 10.0
10:00 235.1 F 8.4
12:00 240.4 F 10.5
14:00 248.6 F 14.8
16:00 261.3 F 22.1
18:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.2
2:00 272.1 F 29.4
3:00 272.1 F 29.0
4:00 272.2 F 29.6

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 218.5 F 0.0
10:00 234.3 F 0.0
15:00 228.5 F 0.0
20:00 226.5 F 0.0
25:00 221.6 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 18:21
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

08/01/2023 11:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____