
Midmark M11 - v1.0.0
Total Cycles: 2089

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2022 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 176.1 F 3.1
4:00 186.7 F 7.2
6:00 210.6 F 12.6
8:00 225.0 F 7.5
10:00 236.8 F 12.8
12:00 236.8 F 9.7
14:00 236.8 F 8.5
16:00 248.8 F 14.5
18:00 261.0 F 21.4
20:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.8 F 29.1
2:00 272.1 F 29.1
3:00 272.6 F 28.9
4:00 272.1 F 29.2

Min 271.3 F 28.6
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 231.7 F 0.0
10:00 240.7 F 0.0
15:00 231.5 F 0.0
20:00 229.4 F 0.0
25:00 226.7 F 0.0
30:00 236.3 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

07/05/2022 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____