
Midmark M11 - v1.0.0
Total Cycles: 3151

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.0 F 0.0
2:00 211.0 F 0.0
4:00 218.6 F 3.1
6:00 227.1 F 8.2
8:00 238.4 F 13.3
10:00 240.0 F 10.7
12:00 236.7 F 8.8
14:00 239.2 F 9.4
16:00 250.5 F 15.2
18:00 262.9 F 22.4
20:00 272.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.7 F 28.8
4:00 272.1 F 28.3

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 233.3 F 0.0
10:00 239.2 F 0.0
15:00 209.0 F 0.0
20:00 203.6 F 0.0
25:00 200.0 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

04/24/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____