
Midmark M11 - v1.0.5
Total Cycles: 2282

Name: RITTER2
Sterilizer ID: V2313228
04/21/2023 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.1 F 0.0
2:00 153.0 F 2.7
4:00 162.5 F 6.0
6:00 187.9 F 10.5
8:00 215.5 F 4.9
10:00 227.3 F 10.1
12:00 230.3 F 7.8
14:00 241.6 F 13.2
16:00 242.3 F 12.0
18:00 255.1 F 18.7
20:00 266.7 F 25.8
22:00 273.3 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.9 F 30.1
2:00 272.0 F 29.5
3:00 272.0 F 30.0
4:00 273.2 F 30.0

Min 271.5 F 29.3
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 209.3 F 0.0
10:00 201.4 F 0.0
15:00 198.7 F 0.0
20:00 198.9 F 0.0
25:00 198.7 F 0.0
30:00 203.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

04/21/2023 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____