

-----  
Midmark M11 - v1.0.0  
Total Cycles: 619

Name: MIDMARK  
Sterilizer ID: V2178198  
01/21/2021 02:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.9 F 0.0  
2:00 134.1 F 1.2  
4:00 160.3 F 3.6  
6:00 185.1 F 7.2  
8:00 204.9 F 11.7  
10:00 217.8 F 5.3  
12:00 230.0 F 10.1  
14:00 232.6 F 8.2  
16:00 241.1 F 12.2  
18:00 239.2 F 10.4  
20:00 252.1 F 16.8  
22:00 263.4 F 23.3  
24:00 271.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.3 F 29.0  
2:00 272.3 F 29.5  
3:00 271.5 F 28.6  
4:00 272.5 F 29.3

Min 271.3 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 225.1 F 0.0  
10:00 230.5 F 0.0  
15:00 226.9 F 0.0  
20:00 228.2 F 0.0  
25:00 227.0 F 0.0  
30:00 237.8 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:02:31  
CYCLE COMPLETE

01/21/2021 03:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_