

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1558

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2022 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.7 F 0.0  
2:00 151.9 F 0.0  
4:00 157.9 F 0.2  
6:00 191.0 F 2.1  
8:00 209.6 F 6.5  
10:00 230.1 F 11.3  
12:00 231.5 F 8.0  
14:00 243.2 F 12.5  
16:00 243.4 F 12.4  
18:00 243.0 F 11.7  
20:00 256.3 F 18.7  
22:00 268.0 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.5 F 29.1  
2:00 271.0 F 28.4  
3:00 273.1 F 29.7  
4:00 271.9 F 28.9

Min 271.0 F 28.4  
Max 273.1 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
5:00 216.8 F 0.0  
10:00 228.0 F 0.0  
15:00 233.6 F 0.0  
20:00 225.9 F 0.0  
25:00 218.1 F 0.0  
30:00 217.2 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 23:17  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:21  
CYCLE COMPLETE

10/04/2022 09:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_