
Midmark M11 - v1.0.0
Total Cycles: 4021

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 178.0 F 2.1
4:00 182.0 F 6.1
6:00 194.6 F 10.8
8:00 215.5 F 3.8
10:00 224.3 F 9.3
12:00 235.7 F 9.4
14:00 240.5 F 11.5
16:00 238.4 F 9.7
18:00 251.9 F 16.2
20:00 263.3 F 22.9
22:00 271.9 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.0
2:00 271.8 F 28.4
3:00 271.5 F 28.6
4:00 272.4 F 29.0

Min 271.1 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 215.0 F 0.0
10:00 218.9 F 0.0
15:00 226.9 F 0.0
20:00 226.1 F 0.0
25:00 225.8 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:52
CYCLE COMPLETE

11/15/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____