

-----  
Midmark M11 - v1.0.5  
Total Cycles: 93

Name: RITTER2  
Sterilizer ID: V2313228  
06/14/2021 08:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.7 F 0.0  
2:00 165.5 F 1.4  
4:00 174.5 F 4.8  
6:00 196.4 F 9.3  
8:00 221.2 F 8.3  
10:00 229.1 F 9.3  
12:00 237.9 F 10.2  
14:00 242.5 F 12.8  
16:00 242.6 F 12.1  
18:00 255.0 F 18.4  
20:00 266.0 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.9 F 29.6  
2:00 271.9 F 29.4  
3:00 272.3 F 29.2  
4:00 272.5 F 29.8

Min 271.6 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 227.9 F 0.0  
10:00 225.6 F 0.0  
15:00 219.2 F 0.0  
20:00 213.7 F 0.0  
25:00 212.7 F 0.0  
30:00 204.3 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 21:33  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:49  
CYCLE COMPLETE

06/14/2021 09:44 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01