
Midmark M11 - v1.0.0
Total Cycles: 4871

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2024 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 73.9 F 0.0
4:00 80.3 F 0.0
6:00 113.3 F 0.1
8:00 147.0 F 0.5
10:00 175.1 F 0.9
12:00 200.0 F 3.3
14:00 211.4 F 6.6
16:00 222.6 F 10.3
18:00 227.6 F 8.6
20:00 234.4 F 9.4
22:00 238.4 F 9.9
24:00 244.1 F 12.4
26:00 241.7 F 11.0
28:00 255.0 F 17.6
30:00 266.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.3 F 28.3
2:00 272.3 F 28.7
3:00 271.8 F 27.9
4:00 272.7 F 28.7

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 208.5 F 0.0
10:00 193.4 F 0.0
15:00 225.2 F 0.0
20:00 227.2 F 0.0
25:00 230.2 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 31:35
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

09/06/2024 10:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____