

-----  
Midmark M11 - v1.0.5  
Total Cycles: 475

Name: RITTER2  
Sterilizer ID: V2313228  
10/22/2021 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.7 F 0.0  
2:00 136.9 F 2.0  
4:00 163.0 F 5.4  
6:00 185.6 F 9.9  
8:00 208.7 F 5.3  
10:00 222.8 F 7.8  
12:00 233.3 F 12.8  
14:00 236.2 F 10.0  
16:00 236.4 F 8.9  
18:00 249.0 F 15.2  
20:00 260.9 F 21.9  
22:00 271.9 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.5 F 29.7  
2:00 271.7 F 29.0  
3:00 271.7 F 29.5  
4:00 272.4 F 29.6

Min 271.1 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 227.9 F 0.0  
10:00 201.2 F 0.0  
15:00 197.3 F 0.0  
20:00 192.9 F 0.0  
25:00 189.8 F 0.0  
30:00 187.1 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 22:35  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:01:04  
CYCLE COMPLETE

10/22/2021 02:11 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02