
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008241
11:00 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:50
STERILIZING 4:50
136.4 C 222kPa 4:50
136.3 C 224kPa 5:20
136.1 C 222kPa 5:50
136.3 C 224kPa 6:20
136.4 C 224kPa 6:50
136.1 C 221kPa 7:20
136.0 C 221kPa 7:50
136.1 C 223kPa 8:20
136.1 C 223kPa 8:20
VENTING 8:20
AIR DRYING 9:13
CYCLE COMPLETE 9:50

Digital Signature #
6452B18B5E3145AD
