
Midmark M11 - v1.0.0
Total Cycles: 4014

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2023 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.6 F 0.0
2:00 173.1 F 2.8
4:00 174.9 F 6.6
6:00 197.1 F 11.0
8:00 215.9 F 4.1
10:00 223.8 F 9.0
12:00 235.8 F 11.2
14:00 237.8 F 10.8
16:00 236.0 F 8.8
18:00 249.6 F 15.3
20:00 260.9 F 21.6
22:00 271.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.1
2:00 271.5 F 28.4
3:00 272.1 F 29.3
4:00 272.1 F 28.9

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 227.3 F 0.0
10:00 233.2 F 0.0
15:00 227.1 F 0.0
20:00 225.9 F 0.0
25:00 226.3 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:40
CYCLE COMPLETE

11/13/2023 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____