
Midmark M11 - v1.0.5
Total Cycles: 2624

Name: RITTER2
Sterilizer ID: V2313228
07/10/2023 08:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.2 F 0.0
2:00 72.2 F 0.0
4:00 101.6 F 0.0
6:00 126.5 F 0.0
8:00 136.4 F 0.5
10:00 161.6 F 3.2
12:00 178.6 F 6.2
14:00 197.3 F 9.5
16:00 216.9 F 13.0
18:00 223.5 F 7.0
20:00 233.5 F 11.2
22:00 231.0 F 7.5
24:00 242.8 F 13.0
26:00 241.2 F 11.4
28:00 253.1 F 17.4
30:00 263.3 F 23.4
32:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 271.8 F 29.2
2:00 272.1 F 29.6
3:00 271.5 F 29.4
4:00 271.6 F 29.0

Min 271.2 F 28.9
Max 272.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 219.4 F 0.0
10:00 230.3 F 0.0
15:00 227.1 F 0.0
20:00 227.7 F 0.0
25:00 229.7 F 0.0
30:00 227.4 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:19
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:10:17
CYCLE COMPLETE

07/10/2023 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____