
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000955
16:38 25/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:29
STERILIZING 4:42
136.0 C 227kPa 4:42
136.4 C 233kPa 5:12
136.4 C 231kPa 5:42
136.4 C 233kPa 6:12
136.2 C 231kPa 6:42
136.4 C 233kPa 7:12
136.4 C 233kPa 7:42
136.3 C 230kPa 8:12
136.3 C 230kPa 8:12
VENTING 8:12
AIR DRYING 9:06
CYCLE COMPLETE 24:06

Digital Signature #
615EE7D8A3344546
