
Midmark M11 - v1.0.0
Total Cycles: 4018

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2023 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.9 F 0.0
2:00 118.4 F 0.8
4:00 118.4 F 3.0
6:00 147.9 F 6.0
8:00 176.5 F 9.6
10:00 200.1 F 7.0
12:00 219.1 F 5.5
14:00 228.4 F 9.4
16:00 238.0 F 11.5
18:00 236.7 F 9.8
20:00 240.7 F 9.2
22:00 246.2 F 13.2
24:00 257.6 F 19.3
26:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.9
2:00 271.7 F 28.2
3:00 271.9 F 28.7
4:00 272.4 F 28.8

Min 271.2 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 216.7 F 0.0
10:00 223.3 F 0.0
15:00 220.5 F 0.0
20:00 231.6 F 0.0
25:00 231.5 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:19
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:05:44
CYCLE COMPLETE

11/14/2023 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____