

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007143  
7:04 04/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:43  
PRESSURIZING 1:46  
STERILIZING 2:53  
136.4 C 220kPa 2:53  
136.2 C 219kPa 3:23  
136.1 C 218kPa 3:53  
136.1 C 222kPa 4:23  
136.4 C 221kPa 4:53  
136.0 C 218kPa 5:23  
136.2 C 222kPa 5:53  
136.4 C 222kPa 6:23  
136.4 C 222kPa 6:23  
VENTING 6:23  
AIR DRYING 7:10  
CYCLE COMPLETE 7:45

-----  
Digital Signature #  
86B617D56D51990E  
-----