
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006421
10:47 06/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:05
STERILIZING 2:54
136.4 C 220kPa 2:54
136.4 C 221kPa 3:24
136.3 C 222kPa 3:54
136.3 C 222kPa 4:24
135.9 C 219kPa 4:54
136.2 C 217kPa 5:24
136.3 C 221kPa 5:54
136.4 C 220kPa 6:24
136.4 C 220kPa 6:24
VENTING 6:25
AIR DRYING 7:14
CYCLE COMPLETE 37:14

Digital Signature #
B6F58143FB6E86E1
