
Midmark M11 - v1.0.5
Total Cycles: 2915

Name: RITTER2
Sterilizer ID: V2313228
09/21/2023 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.1 F 0.0
2:00 80.0 F 0.0
4:00 102.6 F 0.0
6:00 139.2 F 0.0
8:00 169.9 F 0.8
10:00 191.5 F 4.5
12:00 213.5 F 8.4
14:00 222.8 F 12.5
16:00 226.3 F 6.6
18:00 236.7 F 11.0
20:00 233.2 F 7.7
22:00 245.3 F 12.5
24:00 245.0 F 12.7
26:00 257.2 F 19.2
28:00 268.1 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.3 F 28.6
2:00 272.8 F 29.1
3:00 271.7 F 28.5
4:00 272.4 F 28.7

Min 271.6 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 212.3 F 0.0
10:00 195.3 F 0.0
15:00 193.5 F 0.0
20:00 194.0 F 0.0
25:00 191.9 F 0.0
30:00 200.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:14
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:07:15
CYCLE COMPLETE

09/21/2023 03:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____