

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1427

Name: RITTER 1  
Sterilizer ID: V2178198  
12/06/2021 03:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.3 F 0.0  
2:00 176.8 F 3.1  
4:00 199.8 F 7.6  
6:00 217.6 F 12.9  
8:00 226.0 F 7.7  
10:00 238.9 F 12.8  
12:00 238.6 F 10.5  
14:00 236.1 F 9.0  
16:00 251.7 F 16.3  
18:00 264.4 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.8 F 29.3  
2:00 271.5 F 28.4  
3:00 272.7 F 29.2  
4:00 271.6 F 28.4  
  
Min 271.4 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 229.5 F 0.0  
10:00 227.2 F 0.0  
15:00 231.1 F 0.0  
20:00 229.6 F 0.0  
25:00 228.1 F 0.0  
30:00 225.7 F 0.0  
  
Drying Complete

FILLING: 2:27  
HEATING: 19:43  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

12/06/2021 04:39 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02