
Midmark M11 - v1.0.0
Total Cycles: 3872

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.8 F 0.0
2:00 181.1 F 3.7
4:00 199.7 F 7.8
6:00 208.9 F 12.9
8:00 226.7 F 7.2
10:00 236.1 F 12.2
12:00 236.9 F 9.7
14:00 237.2 F 9.2
16:00 249.0 F 14.7
18:00 260.7 F 21.3
20:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.7 F 28.3
2:00 271.8 F 28.6
3:00 272.3 F 28.9
4:00 271.2 F 28.7

Min 271.0 F 28.3
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 207.4 F 0.0
10:00 227.3 F 0.0
15:00 230.0 F 0.0
20:00 228.8 F 0.0
25:00 230.2 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

10/05/2023 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____