
Midmark M11 - v1.0.0
Total Cycles: 2524

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 07:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.0 F 0.0
2:00 179.3 F 4.1
4:00 188.1 F 8.8
6:00 217.0 F 8.0
8:00 226.9 F 9.1
10:00 233.4 F 9.1
12:00 242.5 F 12.3
14:00 241.2 F 10.9
16:00 255.0 F 18.0
18:00 267.0 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.1 F 28.6
2:00 272.3 F 29.1
3:00 271.3 F 28.4
4:00 272.0 F 28.8

Min 271.3 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 226.3 F 0.0
10:00 239.7 F 0.0
15:00 229.5 F 0.0
20:00 228.8 F 0.0
25:00 241.6 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

11/07/2022 08:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____