
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000521
13:32 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 4:53
136.4 C 225kPa 4:53
136.3 C 225kPa 5:23
136.4 C 223kPa 5:53
136.4 C 225kPa 6:23
136.4 C 225kPa 6:53
136.2 C 222kPa 7:23
136.4 C 222kPa 7:53
136.4 C 224kPa 8:23
136.4 C 224kPa 8:23
VENTING 8:23
AIR DRYING 9:16
CYCLE COMPLETE 9:18

Digital Signature #
F5B5E25E5858756C
