
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001421
11:08 14/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:03

Digital Signature #
392ECFD46B721AD3
