
Midmark M11 - v1.0.0
Total Cycles: 693

Name: MIDMARK
Sterilizer ID: V2178198
02/23/2021 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 173.4 F 2.1
4:00 187.9 F 6.4
6:00 211.1 F 11.7
8:00 223.2 F 7.0
10:00 237.7 F 12.9
12:00 238.2 F 10.5
14:00 236.5 F 9.0
16:00 251.3 F 16.2
18:00 263.5 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.1 F 28.9
2:00 273.0 F 29.4
3:00 271.5 F 28.6
4:00 272.6 F 29.2

Min 271.4 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.8 F 0.0
5:00 234.7 F 0.0
10:00 225.4 F 0.0
15:00 228.3 F 0.0
20:00 204.8 F 0.0
25:00 198.0 F 0.0
30:00 198.1 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

02/23/2021 04:53 PM
ACCEPTED: X
REJECTED:
Operator: U01