
Midmark M11 - v1.0.5
Total Cycles: 1058

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 180.4 F 3.0
4:00 193.6 F 7.9
6:00 213.1 F 12.6
8:00 228.1 F 9.4
10:00 234.0 F 9.1
12:00 243.3 F 13.0
14:00 244.1 F 13.0
16:00 257.2 F 20.0
18:00 268.7 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.8 F 30.1
2:00 271.9 F 29.5
3:00 271.4 F 29.5
4:00 272.4 F 30.1

Min 271.2 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 231.3 F 0.0
10:00 224.2 F 0.0
15:00 225.8 F 0.0
20:00 220.8 F 0.0
25:00 220.7 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

05/02/2022 04:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____