
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076

Unit # : 000

WATER QUALITY

2.0uS / 1.2ppm

CYCLE NUMBER 000029

10:52 10/08/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 2:50

PRESSURIZING 7:43

STERILIZING 8:22

123.4 C 119kPa 8:22

123.3 C 120kPa 8:52

123.5 C 120kPa 9:22

123.4 C 120kPa 9:52

123.5 C 121kPa 10:22

123.3 C 120kPa 10:52

123.4 C 120kPa 11:22

123.5 C 120kPa 11:52

123.4 C 120kPa 12:22

123.3 C 120kPa 12:52

123.3 C 120kPa 13:22

123.4 C 120kPa 13:52

123.2 C 119kPa 14:22

123.3 C 119kPa 14:53

123.2 C 119kPa 15:23

123.3 C 120kPa 15:53

123.0 C 118kPa 16:23

123.2 C 118kPa 16:53

123.4 C 120kPa 17:23

123.2 C 118kPa 17:53

123.3 C 119kPa 18:23

123.5 C 120kPa 18:53

123.4 C 120kPa 19:23

123.4 C 120kPa 19:53

123.5 C 120kPa 20:23

123.0 C 117kPa 20:53

123.4 C 120kPa 21:23

123.5 C 120kPa 21:53

123.4 C 120kPa 22:23

123.4 C 120kPa 22:53

123.5 C 120kPa 23:22

VENTING 23:22

AIR DRYING 24:09
CYCLE COMPLETE 54:09

Digital Signature #
4E5E34C801FC3215
