
Midmark M11 - v1.0.0
Total Cycles: 4181

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2024 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.7 F 0.0
2:00 150.4 F 2.4
4:00 168.8 F 5.6
6:00 198.6 F 9.9
8:00 214.5 F 5.9
10:00 224.2 F 7.8
12:00 234.2 F 12.3
14:00 235.8 F 8.7
16:00 242.3 F 11.0
18:00 245.3 F 12.7
20:00 256.8 F 18.9
22:00 267.3 F 25.3
24:00 272.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.8 F 29.1
2:00 271.7 F 28.9
3:00 272.0 F 28.7
4:00 272.5 F 29.2

Min 271.2 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 231.8 F 0.0
10:00 240.4 F 0.0
15:00 228.8 F 0.0
20:00 227.3 F 0.0
25:00 217.2 F 0.0
30:00 214.9 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

01/04/2024 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____