
Midmark M11 - v1.0.5
Total Cycles: 4465

Name: RITTER2
Sterilizer ID: V2313228
03/27/2025 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.2 F 0.0
2:00 166.3 F 2.2
4:00 171.8 F 5.9
6:00 193.5 F 10.2
8:00 220.0 F 6.1
10:00 221.2 F 9.6
12:00 233.6 F 10.1
14:00 238.9 F 12.1
16:00 239.3 F 10.8
18:00 251.7 F 17.1
20:00 262.4 F 23.6
22:00 272.5 F 30.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.9
1:00 271.9 F 30.5
2:00 271.3 F 29.9
3:00 271.5 F 30.6
4:00 272.0 F 30.3

Min 271.1 F 29.8
Max 272.6 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 186.8 F 0.0
10:00 188.0 F 0.0
15:00 190.5 F 0.0
20:00 188.3 F 0.0
25:00 186.4 F 0.0
30:00 188.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

03/27/2025 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____