
Midmark M11 - v1.0.5
Total Cycles: 836

Name: RITTER2
Sterilizer ID: V2313228
02/24/2022 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.7 F 0.0
2:00 168.4 F 3.2
4:00 186.8 F 7.9
6:00 208.9 F 13.5
8:00 224.8 F 8.5
10:00 235.7 F 10.7
12:00 241.4 F 12.5
14:00 242.2 F 11.8
16:00 255.9 F 19.0
18:00 267.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.8 F 29.6
2:00 271.6 F 28.9
3:00 272.8 F 29.8
4:00 272.2 F 29.2

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 196.2 F 0.0
10:00 221.9 F 0.0
15:00 226.8 F 0.0
20:00 222.7 F 0.0
25:00 222.9 F 0.0
30:00 216.3 F 0.0

Drying Complete

FILLING: 1:52
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

02/24/2022 12:41 PM
ACCEPTED: X
REJECTED:
Operator: U02