
Midmark M11 - v1.0.5
Total Cycles: 281

Name: RITTER2
Sterilizer ID: V2313228
08/18/2021 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 81.4 F 0.0
4:00 127.2 F 0.1
6:00 120.9 F 0.1
8:00 146.7 F 1.7
10:00 194.4 F 4.8
12:00 211.3 F 8.2
14:00 224.1 F 12.0
16:00 222.1 F 6.3
18:00 233.1 F 10.5
20:00 231.7 F 8.6
22:00 239.3 F 12.2
24:00 239.2 F 10.3
26:00 251.1 F 16.3
28:00 261.7 F 22.3
30:00 270.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.7
1:00 272.5 F 30.0
2:00 272.4 F 29.7
3:00 272.0 F 29.9
4:00 272.4 F 30.1

Min 271.8 F 29.6
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 236.9 F 0.0
10:00 219.8 F 0.0
15:00 218.1 F 0.0
20:00 205.4 F 0.0
25:00 221.1 F 0.0
30:00 212.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:17
CYCLE COMPLETE

08/18/2021 09:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____