

-----  
Midmark M11 - v1.0.5  
Total Cycles: 936

Name: RITTER2  
Sterilizer ID: V2313228  
03/31/2022 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
2:00 195.3 F 1.5  
4:00 202.9 F 6.8  
6:00 219.7 F 12.4  
8:00 229.8 F 9.1  
10:00 236.4 F 9.7  
12:00 242.5 F 12.1  
14:00 245.4 F 13.8  
16:00 258.6 F 21.1  
18:00 270.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.6  
1:00 272.9 F 30.2  
2:00 271.8 F 29.6  
3:00 273.0 F 30.4  
4:00 272.3 F 29.7  
  
Min 271.8 F 29.5  
Max 273.1 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 219.9 F 0.0  
10:00 227.1 F 0.0  
15:00 229.7 F 0.0  
20:00 226.6 F 0.0  
25:00 225.3 F 0.0  
30:00 227.1 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 18:49  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:43  
CYCLE COMPLETE

03/31/2022 11:04 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_