

-----  
Midmark M11 - v1.0.0  
Total Cycles: 869

Name: RITTER 1  
Sterilizer ID: V2178198  
06/08/2021 06:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 83.7 F 0.0  
4:00 79.7 F 0.0  
6:00 109.0 F 0.1  
8:00 143.5 F 0.1  
10:00 169.8 F 0.2  
12:00 189.9 F 0.3  
14:00 212.9 F 2.9  
16:00 216.2 F 5.7  
18:00 225.1 F 8.5  
20:00 233.2 F 11.9  
22:00 232.3 F 7.4  
24:00 242.9 F 12.1  
26:00 239.4 F 9.9  
28:00 239.3 F 9.5  
30:00 248.7 F 14.2  
32:00 259.5 F 20.1  
34:00 269.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.6 F 28.7  
2:00 272.1 F 28.2  
3:00 271.5 F 27.8  
4:00 272.2 F 28.5  
  
Min 271.4 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
5:00 194.1 F 0.0  
10:00 191.7 F 0.0  
15:00 191.4 F 0.0  
20:00 190.9 F 0.0  
25:00 190.4 F 0.0  
30:00 194.2 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 35:12  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:13:31  
CYCLE COMPLETE

06/08/2021 08:04 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_