
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 003728
15:39 30/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:39
STERILIZING 3:41
136.0 C 220kPa 3:41
136.3 C 222kPa 4:11
136.2 C 225kPa 4:41
136.1 C 223kPa 5:11
136.4 C 225kPa 5:41
136.4 C 225kPa 6:11
136.4 C 226kPa 6:41
136.4 C 224kPa 7:11
136.4 C 224kPa 7:11
VENTING 7:11
AIR DRYING 8:04
CYCLE COMPLETE 8:16

Digital Signature #
773B13C63110E26D
