

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1437

Name: RITTER2  
Sterilizer ID: V2313228  
08/31/2022 09:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 81.4 F 0.0  
2:00 83.9 F 0.0  
4:00 92.2 F 0.0  
6:00 124.9 F 0.7  
8:00 153.7 F 2.4  
10:00 191.4 F 4.5  
12:00 200.1 F 7.6  
14:00 213.1 F 11.1  
16:00 226.1 F 6.0  
18:00 231.0 F 9.0  
20:00 239.8 F 13.1  
22:00 236.9 F 10.1  
24:00 238.3 F 9.2  
26:00 245.9 F 13.6  
28:00 256.4 F 19.2  
30:00 265.9 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.2 F 29.7  
2:00 272.3 F 29.3  
3:00 271.4 F 28.7  
4:00 271.7 F 29.3  
  
Min 271.2 F 28.7  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 203.2 F 0.0  
10:00 213.5 F 0.0  
15:00 215.3 F 0.0  
20:00 217.1 F 0.0  
25:00 217.5 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:58  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:09:52  
CYCLE COMPLETE

08/31/2022 10:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_