
Midmark M11 - v1.0.0
Total Cycles: 4597

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2024 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.7 F 0.0
2:00 111.7 F 0.9
4:00 126.7 F 3.2
6:00 155.6 F 5.8
8:00 187.8 F 9.0
10:00 206.6 F 13.2
12:00 213.7 F 4.7
14:00 220.7 F 8.5
16:00 231.9 F 12.7
18:00 233.3 F 8.8
20:00 241.5 F 12.1
22:00 242.7 F 12.0
24:00 254.1 F 17.8
26:00 264.6 F 23.9
28:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 272.2 F 29.4
2:00 272.0 F 28.9
3:00 271.1 F 28.4
4:00 271.6 F 29.1

Min 271.0 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 211.4 F 0.0
10:00 202.2 F 0.0
15:00 213.9 F 0.0
20:00 235.3 F 0.0
25:00 237.0 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:08
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:45
CYCLE COMPLETE

05/30/2024 01:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____