
Midmark M11 - v1.0.5
Total Cycles: 549

Name: RITTER2
Sterilizer ID: V2313228
11/15/2021 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.1 F 0.0
2:00 189.9 F 3.4
4:00 205.2 F 9.3
6:00 223.7 F 7.3
8:00 237.6 F 11.2
10:00 235.1 F 9.2
12:00 238.2 F 9.6
14:00 249.6 F 15.4
16:00 261.9 F 22.5
18:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.5
2:00 271.4 F 28.8
3:00 272.7 F 29.7
4:00 272.0 F 28.9

Min 271.3 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 234.9 F 0.0
10:00 228.8 F 0.0
15:00 230.3 F 0.0
20:00 225.6 F 0.0
25:00 232.8 F 0.0
30:00 217.2 F 0.0

Drying Complete

FILLING: 2:32
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

11/15/2021 11:26 AM
ACCEPTED: X
REJECTED:
Operator: U02