
Midmark M11 - v1.0.5
Total Cycles: 1399

Name: RITTER2
Sterilizer ID: V2313228
08/16/2022 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.3 F 0.0
2:00 182.6 F 2.0
4:00 206.8 F 7.5
6:00 224.2 F 13.1
8:00 232.2 F 9.3
10:00 237.9 F 9.8
12:00 245.4 F 13.5
14:00 244.8 F 12.8
16:00 257.4 F 19.6
18:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.8
1:00 272.4 F 29.1
2:00 272.3 F 29.5
3:00 271.5 F 28.8
4:00 272.1 F 29.0

Min 271.5 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 184.9 F 0.0
10:00 152.1 F 0.0
15:00 151.1 F 0.0
20:00 145.5 F 0.0
25:00 144.3 F 0.0
30:00 157.4 F 0.0

Drying Complete

FILLING: 2:33
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

08/16/2022 01:14 PM
ACCEPTED: X
REJECTED:
Operator: U02