

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002368  
7:07 06/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:26  
STERILIZING 3:14  
136.0 C 220kPa 3:14  
136.4 C 221kPa 3:44  
136.4 C 223kPa 4:14  
136.4 C 222kPa 4:44  
136.4 C 217kPa 5:14  
136.4 C 221kPa 5:44  
136.4 C 222kPa 6:14  
136.4 C 219kPa 6:44  
136.4 C 219kPa 6:44  
VENTING 6:44  
AIR DRYING 7:35  
CYCLE COMPLETE 7:36  
-----

Digital Signature #  
5A1E728F8F2526B0  
-----