

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1782

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2022 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.4 F 0.0  
2:00 79.8 F 0.0  
4:00 83.8 F 0.0  
6:00 107.2 F 0.0  
8:00 163.8 F 0.3  
10:00 185.6 F 0.8  
12:00 202.8 F 4.2  
14:00 214.4 F 7.7  
16:00 226.3 F 11.5  
18:00 229.2 F 7.2  
20:00 240.7 F 12.4  
22:00 239.2 F 10.5  
24:00 236.7 F 9.0  
26:00 251.1 F 15.9  
28:00 262.6 F 22.6  
30:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.3 F 28.8  
2:00 272.1 F 29.2  
3:00 271.7 F 28.5  
4:00 272.5 F 29.3

Min 271.4 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 202.8 F 0.0  
10:00 230.5 F 0.0  
15:00 234.3 F 0.0  
20:00 228.7 F 0.0  
25:00 210.0 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:16  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:12  
CYCLE COMPLETE

11/29/2022 10:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_