
Midmark M11 - v1.0.5
Total Cycles: 690

Name: RITTER2
Sterilizer ID: V2313228
01/12/2022 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.6 F 0.0
2:00 157.5 F 3.0
4:00 176.4 F 7.0
6:00 199.0 F 12.0
8:00 218.2 F 6.5
10:00 231.6 F 11.7
12:00 234.0 F 9.3
14:00 237.8 F 9.7
16:00 248.1 F 14.9
18:00 260.5 F 21.9
20:00 271.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 29.6
2:00 271.9 F 29.6
3:00 272.3 F 29.5
4:00 271.9 F 29.7

Min 271.5 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 217.1 F 0.0
10:00 222.1 F 0.0
15:00 218.6 F 0.0
20:00 216.6 F 0.0
25:00 216.6 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:07
CYCLE COMPLETE

01/12/2022 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____