
Midmark M11 - v1.0.0
Total Cycles: 1070

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2021 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.9 F 0.0
2:00 95.9 F 0.0
4:00 109.2 F 0.0
6:00 142.3 F 0.2
8:00 178.1 F 1.2
10:00 203.3 F 5.0
12:00 217.7 F 9.0
14:00 228.6 F 13.2
16:00 232.1 F 8.8
18:00 241.4 F 11.2
20:00 243.2 F 11.6
22:00 239.1 F 9.7
24:00 252.5 F 16.2
26:00 263.7 F 22.8
28:00 271.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 27.9
1:00 272.4 F 28.3
2:00 271.9 F 28.5
3:00 271.8 F 28.0
4:00 272.5 F 28.7

Min 271.0 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 227.9 F 0.0
10:00 230.3 F 0.0
15:00 228.1 F 0.0
20:00 231.5 F 0.0
25:00 226.8 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:05
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:05:59
CYCLE COMPLETE

08/04/2021 12:07 PM
ACCEPTED: X
REJECTED:
Operator: U01